

**Disposable Medical Face Mask
Level 2 & Level 3**

SW-0614





SUNWAY (SHENZHEN) PRODUCTS LIMITED

Summary

- **Product Brand:** Sunway Medical, Inc.
- **Product Description:** Disposable Medical Face Mask (non-sterile) Level 2 & Level 3
- **Supplier Ref Number:** SW-0614
- **Product Country of Origin:** China
- **Product Name Submitted:** Disposable Medical Face Mask
- **Name of Reg. Authority:** Intertek / POSI / EN / SGS
- **Product Standards:**
 - Intertek: No. GZHT02337923 ASTM F2100-19
 - POSI: ISO13485 : 2016 Cert
 - EN: No:20R000122MT Ver: DXAO-4144-34
 - SGS: HKTDR2020071080
- **Product Mktg Lic Number:** 914403007992014 4/02/07
- **Product Shelf Life:** 2yr
- **Prod. Intended Use:** Disposable Medical Face Mask

CONTENT

- **Photos of Samples**
- **Packaging**
- **Photos of the Production**
- **Photo of the company**
- **Certificates of the Company**
 - **License**
 - **Medical Device**
 - **Registration**
 - **Certificate**
 - **Management System**
 - **Certification**
 - **Record of Foreign**
 - **Trade Operations**
- **CE Certificate**
- **CE Notification Confirmation**
- **ISO Certificate - ISO13485:2016**
- **Test Reports – Interek, EN, SGS**
- **FDA**



SUNWAY (SHENZHEN) PRODUCTS LIMITED

Photos Of The Sample



Package



合格证CERTIFICATION

【产品名称】一次性使用口罩(非医用)
 【Name】 Disposable Face Mask
 【规格型号】 SW-0614 (17.5x9.5cm)
 【Model】 SW-0614 (17.5x9.5cm)
 【主要成分】 无纺布67.5%、熔喷布32.5%
 【Main components】 67.5%non-woven fabric、 32.5% melt-blown fabric
 【执行标准】 T/CTCA 7-2019
 【Standard】 T/CTCA 7-2019
 【包装规格】 50只/盒
 【Packing Spec.】 50pcs/box
 【有效期】 2年
 【Validity】 2 years
 【生产日期】 2020年7月21日
 【Production Date】 2020.07.21
 【批次号】 K200705Y
 【Production LOT】 K200705Y
 【生产商】 深圳市山而威户外装备科技有限公司
 【Manufacturer】 Sunway (Shenzhen) Products Limited
 【生产地址】 深圳市大鹏新区葵涌街道葵新社区银葵路16号
 【Address】 君轩公司E栋厂房101、201、301东、401、501
 101, 201, east of 301, 401&501, Building E, junxuan Company,
 No.16 Yikui Road, Kuixin Community, Kuichong street,
 Dapeng new district, Shenzhen.



Packaging Information

No.	Unit Package	Qty (Pcs) /Carton	Measurement			N.W. (KGS)	G.W. (KGS)
			Length (cm)	Width (cm)	Height (cm)		
1	50pcs/box	2000	47.00	42.00	35.00	8.00	9.00



合格证CERTIFICATION

【产品名称】一次性使用口罩(非医用)
【Name】 Disposable Face Mask
 【规格型号】SW-0614 (17.5x9.5cm)
【Model】 SW-0614 (17.5x9.5cm)
 【主要成分】无纺布67.5%、熔喷布32.5%
【Main components】 67.5%non-woven fabric、32.5% melt-blown fabric
 【执行标准】T/CTCA 7-2019
【Standard】 T/CTCA 7-2019
 【包装规格】50只/盒
【Packing Spec.】 50pcs/box
 【有效期】2年
【Validity】 2 years
 【生产日期】2020年7月21日
【Production Date】 2020.07.21
 【批次号】K200705Y
【Production LOT】 K200705Y
 【生产商】深圳市山而威户外装备科技有限公司
【Manufacturer】 Sunway (Shenzhen) Products Limited
 【生产地址】深圳市大鹏新区葵涌街道葵新社区银葵路16号
【Address】 君轩公司E栋厂房101、201、301东、401、501
 101, 201, east of 301, 401&501, Building E, junxuan Company,
 No.16 Yikui Road, Kuixin Community, Kuichong street,
 Dapeng new district, Shenzhen.






SUNWAY (SHENZHEN) PRODUCTS LIMITED

Photos Of The Production Workshop





SUNWAY (SHENZHEN) PRODUCTS LIMITED

Photo Of The Company



Certificates of the Company



统一社会信用代码
91440300799201478B

营业执照
(副本)



名称 深圳市山而威户外装备科技有限公司

类型 有限责任公司

成立日期 2007年04月02日

法定代表人 李光曦

住所 深圳市大鹏新区葵涌街道葵新社区银葵路16号君轩公司B栋厂房101、301东、401、501

重要提示

1. 商事主体的经营范围由章程确定。经营范围中属于法律、法规规定应当经批准的项目，取得许可审批文件后方可开展相关经营活动。
2. 商事主体经营范围和许可审批项目等有关企业信用事项及年报信息和其他信用信息，请登录左下角的国家企业信用信息公示系统或扫描右上方的二维码查询。
3. 各类商事主体每年须于成立周年之日起两个月内，向商事登记机关提交上一自然年度的年度报告。企业应当按照《企业信息公示暂行条例》第十条的规定向社会公示企业信息。

登记机关 

2019年09月05日

国家企业信用信息公示系统网址：<http://www.gsxt.gov.cn>

国家市场监督管理总局监制

Certificates of the Company



第二类医疗器械经营备案凭证

备案编号：粤深食药监械经营备 202033323 号

企业名称	深圳市山而威户外装备科技有限公司
法定代表人	李光曦
企业负责人	李光曦
经营方式	批零兼营
住 所	深圳市大鹏新区葵涌街道葵新社区银葵路 16 号君轩公司 E 栋厂房 101、301 东、401、501
经营场所	深圳市大鹏新区葵涌街道葵新社区银葵路 16 号君轩公司 E 栋厂房 101、301 东、401、501
库房地址	深圳市大鹏新区葵涌街道葵新社区银葵路 16 号君轩公司 E 栋厂房 101、301 东、401、501
经营范围	2002 年分类目录（二类）：6801，6802，6803，6804，6805，6806，6807，6808，6809，6810，6812，6813，6815，6816，6820，6821，6822，6823，6824，6825，6826，6827，6828，6830，6831，6832，6833，6834，6840（诊断试剂不需低温冷藏运输贮存），6841，6845，6846，6854，6855，6856，6857，6858，6863，6864，6865，6866，6870，6877，以上类别中包含的植入和介入类产品除外，以上类别中包含的角膜接触镜、助听器产品除外 2017 年分类目录（二类）：01，02，03，04，05，06，07，08，09，10，11，12，13，14，15，16，17，18，19，20，21，22，以上类别中包含的植入和介入类产品除外，以上类别中包含的角膜接触镜、助听器产品除外

备案部门（公章）




备案日期：2020 年 03 月 11 日

Certificates of the Company

医疗器械网络销售信息公示			
关闭			
企业信息			
企业名称	深圳市山而威户外装备科技有限公司	社会信用代码	91440300799201478B
法定代表人	李光耀	企业负责人	李光耀
住所	深圳市大鹏新区葵涌街道葵新社区银葵路16号君轩公司E栋厂房101、301东、401、501	经营场所或生产场所	深圳市大鹏新区葵涌街道葵新社区银葵路16号君轩公司E栋厂房101、301东、401、501
库房地址	深圳市大鹏新区葵涌街道葵新社区银葵路16号君轩公司E栋厂房101、301东、401、501	备案编号	粤深械网备202003110316
医疗器械生产(经营)许可证或备案凭证编号	粤深食药监械经营备202033323号	主体业态(可多选)	<input type="checkbox"/> 医疗器械生产 <input type="checkbox"/> 医疗器械批发 <input type="checkbox"/> 医疗器械零售 <input checked="" type="checkbox"/> 医疗器械批零兼营
*经营范围	2002年分类目录(二类): 6801, 6802, 6803, 6804, 6805, 6806, 6807, 6808, 6809, 6810, 6812, 6813, 6815, 6816, 6820, 6821, 6822, 6823, 6824, 6825, 6826, 6827, 6828, 6830, 6831, 6832, 6833, 6834, 6840 (诊断试剂不需低温冷藏运输贮存); 6841, 6845, 6846, 6854, 6855, 6856, 6857, 6858, 6863, 6864, 6865, 6866, 6870, 6877, 以上类别中包含的植入和介入类产品除外, 以上类别中包含的角膜接触镜、助听器产品除外[br]2017年分类目录(二类): 01, 02, 03, 04, 05, 06, 07, 08, 09, 10, 11, 12, 13, 14, 15, 16, 17, 18, 19, 20, 21, 22, 以上类别中包含的植入和介入类产品除外, 以上类别中包含的角膜接触镜、助听器产品除外		
联系人姓名	李光耀	联系人电话	13695180593
销售类型			
医疗器械网络销售类型	<input type="checkbox"/> 自建类 <input checked="" type="checkbox"/> 入驻类	申请日期:	2020-03-11
网站名称		网络客户端应用程序名	
网站域名		网站IP地址	
电信业务经营许可证编号		非经营性互联网信息服务备案编号	
服务器存放地址		互联网药品信息服务资格证书编号(自建类必填)	
医疗器械网络交易服务第三方平台名称	浙江天猫网络有限公司 浙江淘宝网络有限公司 杭州阿里巴巴广告有限公司 叮当快药科技集团有限公司 深圳市腾讯计算机系统有限公司	医疗器械网络交易服务第三方平台备案凭证编号	(浙) 网械平台备字【2018】第00002号 (浙) 网械平台备字【2018】第00004号 (浙) 网械平台备字【2018】第00001号 (京) 网械平台备字【2018】第00006号 (粤) 网械平台备字【2019】第00003号
变更信息			

Certificates of the Company

INGER CERTIFICATION ASSESSMENT SERVICES



管理体系认证证书

认证编号: 117 20 QU 0122-01 ROM


兹证明 深圳市山而威户外装备科技有限公司
统一社会信用代码 91440300799201478B

注册地址 广东省深圳市大鹏新区葵涌街道葵新社区银葵路 16 号君轩公司 E 栋厂房 101、301 东、401、501
审核地址 广东省深圳市大鹏新区葵涌街道银葵路 16 号君轩公司 E 栋厂房 101、201、301 东、401、501

经现场评审满足: GB/T19001-2016/ISO9001:2015 质量管理体系要求

认证范围 **帐篷的生产和销售及照明器具的生产和外销**


核准:






初次发证: 2020 年 01 月 13 日
有效期至: 2023 年 01 月 12 日

上海英格尔认证有限公司

国家认监委批准号: CNCA-R-2003-117
电话: 400-182-9001/+86 21-51114700
网址: www.icas.org.cn
地址: 上海市徐汇区中山西路2368号
华康大厦801室 200235



关注英格尔微信平台

		
第一次年审	第二次年审	第三次年审

本证书的所有权属上海英格尔认证有限公司, 证书信息及有效性可在国家认监委官方网站 (www.cnca.gov.cn) 上查询, 也可通过登录英格尔官方网站或致电英格尔客户服务部进行查询。本证书在国家规定的各行政、资质许可范围及有效期内使用有效。获证组织必须定期接受年度监督审核并经审核合格此证书方继续有效; 如获证组织未能有效维持以上管理体系, 英格尔有权收回其获证资格。

Certificates of the Company

INGEER CERTIFICATION ASSESSMENT SERVICES



MANAGEMENT SYSTEM CERTIFICATE

Certificate No.: 117 20 GU 0122-01 ROM
 This is to certify the quality management systems of
Sunway (Shenzhen) Products Limited
 Unified Social Credit Code 91440300799201478B

Registered Address 101, East of 301, 401, 501, E Building, Junxuan Company,
 16# Yinkui Road, Kaisin Community, Kuichong Street,
 Dapeng New District, Shenzhen, Guangdong, China
 Auditing Address 101, 201, East of 301, 401, 501, E Building, Junxuan
 Company, 16# Yinkui Road, Kuichong Street, Dapeng
 New District, Shenzhen, Guangdong, China

has been assessed and registered as meeting the requirements of
GB/T19001-2016/ISO9001:2015
 Scope of approval
 Production and Sales of Tents, Production and Export of Lighting Fixtures

Signed by:  

First Certification: 13 Jan. 2020
 Expiry Date: 12 Jan. 2023

Shanghai Ingeer Certification Assessment Co., Ltd.
 Certification and Accreditation Administration of PRC (CMAA) 2005.111
 Tel: 400-162-3001/96 21-51154700
 Web: www.ingeer.org.cn
 Add: Room 101, Building 100000, 20000 West Zhongshan Rd,
 Nuhui District, Shanghai, China, 200235






The ownership of this certificate belongs to Shanghai Ingeer Certification Assessment Co., Ltd. The information & validation of this certificate can be checked on the CMAA website: WWW.CMAA.GOV.CN and ICAS website, or by calling ICAS's clients services Dept. This certificate is only valid when used together with related records when appropriate. If the organization fails to effectively maintain the above management system, ICAS will take the right to withdraw the qualification certificate.

Certificates of the Company

对外贸易经营者备案登记表

统一社会信用代码: 91440300799201478B
进出口企业代码: _____

备案登记表编号: 02520799

经营者中文名称	深圳市山而威户外装备科技有限公司		
经营者英文名称	SUNWAY(SHENZHEN) PRODUCTS LIMITED		
组织机构代码	_____	经营者类型 (由备案登记机关填写)	私营有限责任公司
住 所	深圳市大鹏新区葵涌街道葵新社区银葵路16号君轩公司E栋厂房101、301东、401、501		
经营场所 (中文)	深圳市大鹏新区葵涌街道葵新社区银葵路16号君轩公司E栋厂房101、301东、401、501		
经营场所 (英文)	101, east of 301, 401&501, Building E, Junxuan company, NO.16 Yinkui Road, Kuixin Community, Kuichong Street, Dapeng New District, Shenzhen		
联系电话	0755-29450963	联系传真	0755-29451163
邮政编码	518104	电子邮箱	13392872595@163.com
工商登记注册日期	2007-4-2	工商登记注册号	_____

依法办理工商登记的企业还须填写以下内容


企业法定代表人姓名	李光曦	有效证件号	430219197112261639
注册资金	伍佰万元		(折美元)

依法办理工商登记的外国 (地区) 企业或个体工商户 (独资经营者) 还须填写以下内容

企业法定代表人 / 个体工商户负责人姓名		有效证件号	
企业资产 / 个人财产			(折美元)

备注	
----	--

填表前请认真阅读背面的条款, 并由企业法定代表人或个体工商户负责人签字、盖章。



2019 年 月 日

CE Certificate



Certificate of Compliance

93/42/EU Protective Equipment

Certificate Number: TXB20200310SY01
Applicant: Sunway (shenzhen) Products Limited.
Address of Applicant: 101 east of 301,401&501,Building E junxuan company,No .16 yinkui Road, Kuixin Community,Kuichong street,dapeng new District,shenzhen
Manufacturer: Sunway (shenzhen) Products Limited,
Address of Manufacturer: 101, 301 East, 401, 501,E Building , Junxuan Company, 16# Yinkui Road, Kuixin Community, Kui Chong Street, Dapeng District, Shenzhen,518119, China
Product Name: Disposable Medical Face Mask
Trademark: SUN-V
Main Test Model: SW-0614
Sufficient samples of the product have been tested and found to be in conformity with

Test Standard: EN 14683:2005

As Shown in the
Test Report Number: TXB20200310SY01

Conclusion
This Verification of Compliance has been issued on a voluntary basis.. TXB confirms that a Technical Construction File ((TCF)) is existent for the above listed product((s)). The TCF satisfactorily covers the essential requirements off the above listed Directive((s)). Other relevant Directives have to be observed in case they are applicable.. This Document is only valid for the equipment and configuration described and in conjunction with the TCF detailed above.. Whereas the Manufacturer is responsible off the certification off the product((s)) and not exempted to perform all the necessary activities before placing the product((s)) on the market.. The Manufacturer is also responsible off the internal production control to ensure the product((s)) are in compliance with the essential requirements off the above mentioned Directive((s))

Issued date: Mar,10,2020

The certificate applies to the tested sample above-mentioned only and shall not imply an assumption of the conformity of the production. It is only valid in connection with the test report above-mentioned. Copyright of this certificate is reserved by TXB and may not be reproduced other than in full and with the prior approval of the General Manager.

TXB Rheinland Testing Services Corp Limited
E-mail: lacy@xb-lab.com www.xb-lab.com

CE Notification Confirmation



CE Notification Confirmation

This is to confirm that, according to the council directive 93/42/EEC (MDD), SUNGO Europe B.V. performed the notification duties and responsibilities as the European authorized representative of:

SUNWAY (SHENZHEN) PRODUCTS LIMITED
101, East of 301,401&501, Building E, Junxuan Company, No.16
Yinkui Road, Kuixin Community, Kuichong Street, Dapeng New
District, Shenzhen, 518119, China

The Manufacturer has provided SUNGO Europe B.V. with the EC Declaration of Conformity confirming that the medical device, as stipulated here below, is fulfilling the applicable requirements of the European Council Directive 93/42/EEC.

According to 93/42/EEC (MDD), the European Databank on Medical Devices (EUDAMED) is established as of May 1 2011. The Netherlands Competent Authority is notified of the manufacturer's medical devices and has allocated registration number.

Disposable Medical Face Mask
Class I according to Annex IX of 93/42/EEC
GMDN: 35177
CIBG Number: NL-CA002-2020-50600

Where the manufacturer affix's the CE marking to the product listed they must ensure that all the requirements of the appropriate EU directive(s) have and continue to be met.

This document should be used together with the competence authority notification letter and the Declaration of Conformity issued by the Manufacturer. This document will become to be invalid once the notification status is changed or the EAR agreement is terminated.

Reference Number: EUCAN00238
Issue date: May 08, 2020

SUNGO Europe B.V.
Olympisch Stadion 24, 1076DE
Amsterdam, Netherlands
ec.rep@sungogroup.com





POSI

CERTIFICATE

This is to certify that the Quality Management System of

SUNWAY (SHENZHEN) PRODUCTS LIMITED

Business license number: 91440300799201478B

Registered Address: 101, 201, East of 301, 401 & 501, Building E, Junxuan Company, No.16 Yinkui Road, Kuixin Community, Kuichong Street, Dapeng New District, Shenzhen, Guangdong Province, China

Audit Address: 101, 201, East of 301, 401 & 501, Building E, Junxuan Company, No.16 Yinkui Road, Kuixin Community, Kuichong Street, Dapeng New District, Shenzhen, Guangdong Province, China

applicable to

Production and sales of disposable medical isolation clothing, medical isolation face shield, medical isolation eye mask, disposable medical coverall isolation clothing, medical isolation shoe cover(in the filing certificate); Production and sales of daily protective face masks; Production and sales of disposable protective clothing(export only)

has been assessed and registered by POSI against the provisions of

ISO13485:2016

This registration is subject to the company maintaining a quality management system, to the above standard, which will be monitored by POSI.

Please consult the website: www.posicert.com

The certificate information is also available on the CNCA official website: <http://cx.cnca.cn>.



General Manager

Certificate Registration No: 381200045R0M

Initial issue date : 2020.06.11 Issue date : 2020.06.11 Valid until: 2023.06.10



Shanghai POSI Certification Co., Ltd.

Room 1002A, No.1500, Century Avenue, Pudong New Area, Shanghai ,China.Email:info@posicert.com

Test Report (Electronic version)

Verification Website: www.gtcc.net.cn

Verification Code: DXAO-4144-34

EN Test Report

No:20R000122MT

Ver: DXAO-4144-34

No:20R000122MT

Issue Date: 2020-07-20

Applicant: SUNWAY (SHENZHEN) PRODUCTS LIMITED

Address: 101, 201, EAST OF 301, 401&501, BUILDING E, JUNXUAN COMPANY, NO.16 YINKUI ROAD, KUIXIN COMMUNITY, KUICHONG STREET, DAPENG NEW DISTRICT, SHENZHEN, 518119, CHINA

Information confirmed by applicant:

Disposable medical face mask(non-sterile)

Quantity: eighty pieces

Lot number: 20200410

Model: SW-0614

Size: 175mm×95mm

Classification: Type II R

Standard Adopted:

EN 14683:2019+AC:2019 <Medical face masks-Requirements and test methods>

Date Received/Date Test Started: 2020-04-22

Conclusion:

Bacterial filtration efficiency (BFE)	M
Microbial cleanliness	M
Differential pressure	M
Splash resistance pressure	M
Materials and construction	M
Design	M
General	M

Note: "M"-Meet the standard's requirement "F"-Fail to meet the standard's requirement "-"-No comment

Remark:

Modified content: modified applicant address.

This report replaces test report 20R000122MO which has become invalid automatically.

All the tested items are tested under the standard condition (except for indication).

Copies of the report are valid only re-stamped.

The experiment was carried out at No.1, Zhujiang Road, Panyu District, Guangzhou, Guangdong, P.R.China.

Approved By:
ZiShan Guo Senior Engineer

ZiShan Guo



Page 1 of 13

Test Report


(Electronic version)

No: 20R000122

EN Test Report

No:20R000122MT

Ver: DXAO-4144-34



Test Report

(Electronic version)

No: 20R000122

EN Test Report

No:20R000122MT

Ver: DXAO-4144-34

Bacterial filtration efficiency (BFE)

Test method: EN 14683: 2019+AC: 2019 Annex B

Test principle:

A specimen of the mask material is clamped between a six-stage cascade impactor and an aerosol chamber. An aerosol of Staphylococcus aureus is introduced into the aerosol chamber and drawn through the mask material and the impactor under vacuum. The bacterial filtration efficiency (BFE) of the mask is given by the number of colony forming units passing through the medical face mask material expressed as a percentage of the number of colony forming units present in the challenge aerosol.

Test equipment:

Incubator
Electronic balance
Autoclave
Experimental system for bacterial filtration efficiency (BFE) of mask

The environmental conditions of the laboratory and test condition:

Total bacteria: 0 CFU/plate
Total fungi: 0 CFU/plate
Blank experiment: Aseptic growth
Test environment temperature: 24.5℃, Relative humidity: 50.5%
Culture medium: TSA agar medium
Culture temperature: 37℃, Culture time: 48h
Test bacteria : staphylococcus aureus ATCC 6538
Concentration of bacterium: 5.0×10^5 CFU /ml
Positive control average (C): 1.9×10^3 CFU
Negative monitor count: <1 CFU
Test area: 40 cm²
Dimensions of the test specimens: 15cm×15cm
Flow rate: 28.3 l/min
Pretreatment: Condition each specimen for 4 h by exposure to a temperature of (21±5)℃ and a relative humidity of (85±5)%
Mean particle size: 3.0 μm
The medical face mask in contact with the bacterial challenge: inside



Test Report

(Electronic version)

No: 20R000122

EN Test Report

No:20R000122MT

Ver: DXAO-4144-34

Results:

Sample	T	BFE (%)	Requirement (%)	Classification	Conclusion
1	21	98.89	≥98 EN 14683:2019+AC:2019	Type II R	Pass
2	27	98.58			
3	25	98.68			
4	17	99.11			
5	31	98.37			

Remarks:

For each test specimen calculate the bacterial filtration efficiency B, as a percentage, using the following formula:

$$B = (C - T) / C \times 100$$

where

B is bacterial filtration efficiency (BFE), %;

C is positive control average;

T is the total plate count for the test specimen.



Test Report

(Electronic version)

No: 20R000122

EN Test Report

No:20R000122MT

Ver: DXAO-4144-34

Microbial cleanliness

Test method: EN ISO 11737-1:2018, Membrane filtration

Test principle:

Take the required samples from the original packaging. Weigh a certain amount of sample and placed in a sterile 500 ml bottle containing 300 ml of extraction liquid (1 g/l Peptone, 5 g/l NaCl and 2 g/l Tween 20). The bottle is laid down on an orbital shaker and shaken for 5 min at 250 rpm. After this extraction step, 100 ml of the extraction liquid is filtered through a 0.45 μm filter and laid down on a TSA plate for the total viable aerobic microbial count. Another 100 ml aliquot of the same extraction liquid is filtered in the same way and the filter plated on Sabouraud Dextrose agar (SDA) for fungi enumeration. The plates are incubated for 3 days at 30°C and 7 days at (20 to 25)°C for TSA and SDA plates respectively. The total bioburden is expressed by addition of the TSA and SDA counts.

Test equipment:

Constant temperature incubator
Electronic balance
Pressure steam sterilizer
Biosafety cabinet

The environmental conditions of the laboratory and test condition:

Test environment temperature: 24.5°C, Relative humidity: 56.0%

Test environment monitoring: total bacteria: 0 CFU/plate, total fungi: 0 CFU/plate, blank experiment: aseptic growth



Test Report

(Electronic version)

No: 20R000122

EN Test Report

No:20R000122MT

Ver: DXAO-4144-34

Results:

Microbial	Measured value (CFU/g)	Microbial cleanliness (CFU/g)	Requirement (CFU/g)	Classification	Conclusion
Bacteria	25	28	≤30 EN 14683:2019+AC:2019	Type II R	Pass
Fungi	3				



Test Report

(Electronic version)

No: 20R000122

EN Test Report

No:20R000122MT

Ver: DXAO-4144-34

Differential pressure

Test method: EN 14683:2019+AC:2019 Annex C

Test principle:

This procedure was performed to evaluate the differential pressure of the medical face mask material by measuring the air exchange pressure through a measured surface area at a constant air flow rate.

Test equipment:

GTTC-YLC-1 Apparatus for measuring differential pressure

The environmental conditions of the laboratory and test condition:

Air flow: 8 l/min

Test area: 4.9cm²

Pretreatment: Condition each specimen for a minimum of 4 h by exposure to a temperature of (21±5) °C and a relative humidity of (85±5)%

General location of the areas of the mask the differential measurements: specimen center



Test Report

(Electronic version)

No: 20R000122

EN Test Report

No:20R000122MT

Ver: DXAO-4144-34

Results:

Sample	Measured value (Pa)	Differential pressure (Pa/cm ²)	Requirement (Pa/cm ²)	Classification	Conclusion
1	88	18.0	<60 EN 14683:2019+AC:2019	Type II R	Pass
2	82				
3	80				
4	95				
5	93				
Average	88				



Test Report

(Electronic version)

No: 20R000122

EN Test Report

No:20R000122MT

Ver: DXAO-4144-34

Splash resistance pressure

Test method: ISO 22609:2004

Test principle:

A specimen medical face mask is supported on an apparatus. A volume of synthetic blood is sprayed horizontally at the specimen mask to simulate the scenario of a mask being splashed by a punctured blood vessel. The volume of fluid, distance to impact, orifice size and fluid velocity are defined in this method and intended to be consistent with this health care scenario. Any evidence of synthetic blood penetration on the side of the medical face mask contacting the wearer's face constitutes failure. Results are reported as "pass/fail". Specimen medical face masks are evaluated at a total of three different velocities corresponding to human blood pressures of 10.6 kPa, 16.0 kPa, and 21.3 kPa. Test results are reported at each velocity and the medical face mask is rated at the highest corresponding blood pressure for which medical face mask specimens demonstrate an acceptable quality limit of 4.0.

Test equipment:

Test apparatus for synthetic blood penetration LFY-227

Air compressor

Graduated cylinder

Electronic balance

Targeting plate

The environmental conditions of the laboratory and test condition:

Condition each specimen for a minimum of 4 h by exposure to a temperature of $(21\pm5)^{\circ}\text{C}$ and a relative humidity of $(85\pm5)\%$

Surface tension of synthetic blood: 0.042 N/m

Pressure: 16.0 kPa

Velocity: 550 cm/s



Test Report

(Electronic version)

No: 20R000122

EN Test Report

No:20R000122MT

Ver: DXAO-4144-34

Results:

Sample	Measured value	Requirement (kPa)	Classification	Conclusion
	Pressure			
	16.0 kPa			
1	pass	≥16.0 EN 14683:2019+AC:2019	Type II R	Pass
2	pass			
3	pass			
4	pass			
5	pass			
6	pass			
7	pass			
8	pass			
9	pass			
10	pass			
11	pass			
12	pass			
13	pass			
14	pass			
15	pass			
16	pass			
17	pass			
18	pass			
19	pass			
20	pass			
21	pass			
22	pass			
23	pass			
24	pass			
25	pass			
26	pass			
27	pass			
28	pass			
29	pass			
30	pass			
31	pass			
32	pass			
Final result	pass			

Remarks:

An acceptable quality limit of 4.0 % is met for a single sampling plan when 29 or more of the show "pass" results.



Test Report

(Electronic version)

No: 20R000122

EN Test Report

No:20R000122MT

Ver: DXAO-4144-34

Materials and construction

Test Method: EN 14683:2019+AC:2019 5.1.1

Results:

Requirement	Conclusion
The medical face mask is a medical device, generally composed of a filter layer that is placed, bonded or moulded between layers of fabric.	Pass
The medical face mask shall not disintegrate, split or tear during intended use.	Pass
In the selection of the filter and layer materials, attention shall be paid to cleanliness.	Pass



Test Report

(Electronic version)

No: 20R000122

EN Test Report

No:20R000122MT

Ver: DXAO-4144-34

Design

Test Method: EN 14683:2019+AC:2019 5.1.2

Results:

Requirement	Conclusion
The medical face mask shall have a means by which it can be fitted closely over the nose, mouth and chin of the wearer and which ensures that the mask fits closely at the sides.	Pass
Medical face masks may have different shapes and constructions as well as additional features such as a face shield (to protect the wearer against splashes and droplets) with or without anti-fog function, or a nose bridge (to enhance fit by conforming to the nose contours).	Pass



Test Report

(Electronic version)

No: 20R000122

EN Test Report

No:20R000122MT

Ver: DXAO-4144-34

General

Test Method: EN 14683:2019+AC:2019 5.2.1

Results:

Requirement	Conclusion
All tests shall be carried out on finished products or samples cut from finished products.	Pass

——End of Report——



Test Report

No. GZHT02337923
ASTMF2100-19

Report Ref:	GZHT02337923		
Date Received:	Sep 14, 2020	Date Issued:	Oct 09, 2020

Company Name:	SUNWAY (SHENZHEN) PRODUCTS LIMITED
Address:	101, 201, EAST OF 301, 401 &501 BUILDING E, JUNXUAN COMPANY NO.16 YINKUI ROAD, KUIXIN COMMUNITY KUICHONG STREET, DAPENG NEW DISTRICT, SHENZHEN, CHINA
Contact Name:	王勤

The Following Sample Was Submitted And Identified By/On Behalf Of The Applicant As:	
End Uses	: Medical Face Mask
Ratings	: Level 3
Sample Name	: Disposable Medical Face Mask
No. Of Sample	: One(100 pieces)
Size	: -
Colour	: Blue
Standard	: ASTM F2100-19 ^{E1}
Date received/ Test Started	: Sep 14, 2020
Ref	: Type No.: SW-0614

Test was conducted on specific items, at our client's request.

Prepared And Checked By:
For Intertek Testing Services Shenzhen Ltd. Guangzhou Branch



Lin Lin
General Manager

HTJ2337923 Page 1 Of 8

QIN / hilaryxu

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch
深圳天祥质量技术服务有限公司广州分公司

Room 02, 1-8/F & Room 01, E101/E201/E301/E401/E501/E601/E701/E801, Hengyun Building, 235 Kaifa Ave., Guangzhou
No.7-2, Cabin Road, Guangzhou Science City, GETDD, Guangzhou, China Economic & Technological Development District, Guangzhou, China
中国广州经济技术开发区科学城彩频路7号之二第1-8层02房、01房 02楼 01楼
中国广州经济技术开发区开发大道235号恒运大厦3楼
Tel: +86 20 8213 9001 Fax: +86 20 8208 9909 Postcode: 510663
Tel: +86 20 8396 6868 Fax: +86 20 8222 8169 Postcode: 510730



Test Report

No. GZHT02337923

ASTMF2100-19



Prepared And Checked By:
For Intertek Testing Services Shenzhen Ltd. Guangzhou Branch

Lin Lin
General Manager

QIN / hilaryxu

HTJ2337923 Page 2 Of 8

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch
深圳天祥质量技术服务有限公司广州分公司

Room 02, 1-8/F & Room 01, E101/E201/E301/E401/E501/E601/E701/E801, Hengyun Building, 235 Kaifa Ave., Guangzhou
No.7-2, Catbin Road, Guangzhou Science City, GETDD, Guangzhou, China Economic & Technological Development District, Guangzhou, China
中国广州经济技术开发区科学城彩频路 7号之二第 1-8层 02房、01房 中国广州经济技术开发区开发大道 235号恒运大厦 3楼
Tel: +86 20 8213 9001 Fax: +86 20 8208 9909 Postcode: 510663 Tel: +86 20 8396 6868 Fax: +86 20 8222 8169 Postcode: 510730



Test Report

No. GZHT02337923

ASTMF2100-19

Summary of testing:

With reference to following standard:

ASTM F2100-19¹ Standard Specification for Performance of Materials Used in Medical Face Masks Level 3

Materials Used in The Submitted Sample Were Found To Comply With The Level 3 Requirements of ASTM F2100-19¹with respect to Bacterial Filtration Efficiency (BFE), Differential Pressure, Sub-Micron Particulate Filtration, Resistance to Penetration by Synthetic Blood and Flammability tests.

Prepared And Checked By:
For Intertek Testing Services Shenzhen Ltd. Guangzhou Branch



Lin Lin
General Manager

QIN / hilaryxu

HTJ2337923 Page 3 Of 8

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch
深圳天祥质量技术服务有限公司广州分公司

Room 02, 1-8/F & Room 01, E101/E201/E301/E401/E501/E601/E701/E801, Hengyun Building, 235 Kaifa Ave., Guangzhou
No.7-2, Catbin Road, Guangzhou Science City, GETDD, Guangzhou, China Economic & Technological Development District, Guangzhou, China
中国广州经济技术开发区科学城彩频路 7号之二第 1-8层 02房, 恒运大厦, 中国广州经济技术开发区开发大道 235号恒运大厦 3楼
101, E201, E301, E401, E501, E601, E701, E801 Tel: +86 20 8396 6868 Fax: +86 20 8222 8169 Postcode: 510730
Tel: +86 20 8213 9001 Fax: +86 20 8208 9909 Postcode: 510663



Test Report

Number: GZHT02337923

Tests Conducted (As Requested By The Applicant)

Test Report

No. GZHT02337923
ASTMF2100-19

- Differential Pressure (ASTM F2100-19¹, Section 9.2, Testing Refer to EN 14683:2019+AC:2019 Annex C):
Air flow: 8L/min, Test Area Diameter 25 mm, Test Area: 4.9 cm².

Tested Sample	Result (mm H ₂ O/cm ²)*					Performance Requirement for Medical Face Mask (mm H ₂ O/cm ²)
	Specimen 1	Specimen 2	Specimen 3	Specimen 4	Specimen 5	
Location 1	5.4	6.0	5.2	6.1	5.4	Level 3 <6.0
Location 2	5.1	5.6	6.8	6.3	5.4	
Location 3	6.2	5.6	6.6	6.0	6.1	
Location 4	5.9	5.6	6.0	6.3	6.5	
Location 5	6.1	6.2	5.1	5.0	5.8	
Average	5.7	5.8	5.9	5.9	5.8	

* = All the locations were evenly taken from the main mask body.

QIN / hilaryxu

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch
深圳天祥质量技术服务有限公司广州分公司

Room 02, 1-8/F & Room 01, E101/E201/E301/E401/E501/E601/E701/E801, Hengyun Building, 235 Kaifa Ave., Guangzhou
No.7-2, Cainin Road, Guangzhou Science City, GETDD, Guangzhou Economic & Technological Development District, Guangzhou, China
中国广州经济技术开发区科学城彩虹路 7号之二第 1-8层 02房, 恒运大厦, 中国广州经济技术开发区开发大道 235号恒运大厦 3楼
101, E201, E301, E401, E501, E601, E701, E801
Tel: +86 20 8213 9001 Fax: +86 20 8208 9909 Postcode: 510663
Tel: +86 20 8396 6868 Fax: +86 20 8222 8169 Postcode: 510730



Test Report

No. GZHT02337923

ASTMF2100-19

- 2 Resistance to Penetration by Synthetic Blood (ASTM F2100-19, Section 9.4, Testing Refer to ASTM F1862/F1862M-17, Test Pressure: 160 mmHg, Velocity: 635 cm/s):
Condition test specimens for a minimum of 4 hours in an environment of temperature (21±5) °C and relative humidity (85±5)% and conduct the test within 1 minute of removal from conditioning chamber.
Test Environment Condition: Temperature 24.0°C, Relative Humidity 86.0%

Tested Sample	Observation	Pass/Fail	Performance Requirement for Medical Face Mask Level 3: No Penetration at 160 mm Hg
Specimen (1)	No penetration	Pass	
Specimen (2)	No penetration	Pass	
Specimen (3)	No penetration	Pass	
Specimen (4)	No penetration	Pass	
Specimen (5)	No penetration	Pass	
Specimen (6)	No penetration	Pass	
Specimen (7)	No penetration	Pass	
Specimen (8)	No penetration	Pass	
Specimen (9)	No penetration	Pass	
Specimen (10)	No penetration	Pass	
Specimen (11)	No penetration	Pass	
Specimen (12)	No penetration	Pass	
Specimen (13)	No penetration	Pass	
Specimen (14)	No penetration	Pass	
Specimen (15)	No penetration	Pass	
Specimen (16)	No penetration	Pass	
Specimen (17)	No penetration	Pass	
Specimen (18)	No penetration	Pass	
Specimen (19)	No penetration	Pass	
Specimen (20)	No penetration	Pass	
Specimen (21)	No penetration	Pass	
Specimen (22)	No penetration	Pass	
Specimen (23)	No penetration	Pass	
Specimen (24)	No penetration	Pass	

QIN / hilaryxu



Test Report

Number: GZHT02337923

Tests Conducted (As Requested By The Applicant)

Test Report

No. GZHT02337923
ASTMF2100-19

Specimen (25)	No penetration	Pass
Specimen (26)	No penetration	Pass
Specimen (27)	No penetration	Pass
Specimen (28)	No penetration	Pass
Specimen (29)	No penetration	Pass
Specimen (30)	No penetration	Pass
Specimen (31)	No penetration	Pass
Specimen (32)	No penetration	Pass

Conclusion*: **Accepted**

* = An acceptable quality limit of 4.0% is met for a single sampling plan when 29 or more of the 32 tested specimens show "pass" results.

3 Sub-Micron Particulate Filtration (ASTM F2100-19⁹, Section 9.3, Testing Refer to ASTM F2299/F2299M-17):

Particle size in aerosol: 0.1 μm, Aerosol: Polystyrene Latex Spheres (PSL), Test area: 100 cm² Airflow: 5.33 cm/s, Sampling time: 1 min.

Laboratory Condition: Temperature 21.0°C, Relative Humidity 47.0%

<u>Tested Sample/Component</u>	<u>Result (%)</u>	<u>Performance Requirement for Medical Face Mask (%)</u>
Specimen (1)	99.3	Level 3: ≥98
Specimen (2)	99.4	
Specimen (3)	99.5	
Specimen (4)	98.7	
Specimen (5)	98.8	

QIN / hilaryxu

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch
深圳天祥质量技术服务有限公司广州分公司

Room 02, 1-8/F & Room 01, E101/E201/E301/E401/E501/E601/E701/E801, Hengyun Building, 235 Kaifa Ave., Guangzhou
No.7-2, Cainin Road, Guangzhou Science City, GETDD, Guangzhou Economic & Technological Development District, Guangzhou, China
中国广州经济技术开发区科学城彩虹路7号之二第1-8层02房、01房 中国广州经济技术开发区开发大道235号恒运大厦3楼
101, E301, E401, E501, E601, E701, E801 Tel: +86 20 8396 6868 Fax: +86 20 8222 8169 Postcode: 510730
Tel: +86 20 8213 9001 Fax: +86 20 8208 9909 Postcode: 510663



Tests Conducted (As Requested By The Applicant)

Test Report

No. GZHT02337923

ASTMF2100-19

4 Flammability Test (ASTM F2100-19¹, Section 9.5, Testing Refer to 16 CFR Part 1610 (As Amendment In 2008)):

X	Plain Surface	Raised Surface	
	Burn Direction : 6 Length .Width		Requirement Class 1
	Prelim Plain Surface :		
	Length: IBE		
	Width: DNI		
	Original* (seconds)		
1.	IBE		
2.	IBE		
3.	DNI		
4.	DNI		
5.	DNI		
6.	-		
7.	-		
8.	-		
9.	-		
10.	-		
	Average : -		

Classification :	6	Class 1, Normal Flammability
	...	Class 2, Intermediate Flammability, Raised Surface
	...	Class 3, Rapid And Intense Burning

Explanation Of Flammability Results:	
DNI	Did not ignite.
IBE	Ignited but extinguished.

* The disposable fabrics and garments need not to be refurbished in accordance with 16 CFR Part 1610.35 (a)(2).

QIN / hilaryxu

Page 7 Of 8

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch
深圳天祥质量技术服务有限公司广州分公司

Room 02, 1-8/F & Room 01, E101/E201/E301/E401/E501/E601/E701/E801, Hengyun Building, 235 Kaifa Ave., Guangzhou
No.7-2, Cainin Road, Guangzhou Science City, GETDD, Guangzhou, China Economic & Technological Development District, Guangzhou, China
中国广州经济技术开发区科学城彩频路 7号之二第 1-8层 02房, 恒运大厦, 中国广州经济技术开发区开发大道 235号恒运大厦 3楼
101, E201, E301, E401, E501, E601, E701, E801
Tel: +86 20 8213 9001 Fax: +86 20 8208 9909 Postcode: 510663
Tel: +86 20 8396 6868 Fax: +86 20 8222 8169 Postcode: 510730



Tests Conducted (As Requested By The Applicant)

Test Report

No. GZHT02337923

ASTMF2100-19

5 Bacterial Filtration Efficiency (BFE)

As per ASTM F2100-19¹ Standard Specification for Performance of Materials Used in Medical Face Masks Section 9.1 and ASTM F2101-19 Standard Test Method for Evaluating the Bacterial Filtration Efficiency (BFE) of Medical Face Mask Materials, Using a Biological Aerosol of Staphylococcus aureus.

Test Item	Results (%)					Performance Requirement for Medical Face Mask (%)
	Specimen (1)	Specimen (2)	Specimen (3)	Specimen (4)	Specimen (5)	
Bacterial Filtration Efficiency (BFE)	99.9	99.9	99.9	>99.9	>99.9	Level 3: ≥ 98

Remarks:

1. Biological Aerosol: Staphylococcus aureus (ATCC 6538).
2. Testing side: Inside of the test specimen was facing towards the challenge aerosol.
3. Test area: 78 cm²
4. Flow rate: 28.3 L/min
5. The average plate count results of the positive controls: 2.6×10⁸ CFU
6. The average plate count results of the negative controls: < 1 CFU
7. CFU = Colony Forming Unit

This item was conducted in Caipin Road, Guangzhou Science City, GETDD, Guangzhou, Guangdong.

End of Report

This report is made solely on the basis of your instructions and/or information and materials supplied by you. It is not intended to be a recommendation for any particular course of action. Intertek does not accept a duty of care or any other responsibility to any person other than the Client in respect of this report and only accepts liability to the Client insofar as is expressly contained in the terms and conditions governing Intertek's provision of services to you. Intertek makes no warranties or representations either express or implied with respect to this report save as provided for in those terms and conditions. We have aimed to conduct the Review on a diligent and careful basis and we do not accept any liability to you for any loss arising out of or in connection with this report, in contract, tort, by statute or otherwise, except in the event of our gross negligence or wilful misconduct. No copy of the test report(except for full text copy) shall be made without the written approval by Intertek.

QIN / hilaryxu

Page 8 Of 8

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch
深圳天祥质量技术服务有限公司广州分公司

Room 02, 1-8/F & Room 01, E101/E201/E301/E401/E501/E601/E701/E801, Hengyun Building, 235 Kaifa Ave., Guangzhou
No.7-2, Caipin Road, Guangzhou Science City, GETDD, Guangzhou, China Economic & Technological Development District, Guangzhou, China
中国广州经济技术开发区科学城彩频路 7号之二第 1-8层 02房, 恒运大厦, 中国广州经济技术开发区开发大道 235号恒运大厦 3楼
101, E201, E301, E401, E501, E601, E701, E801 Tel: +86 20 8396 6868 Fax: +86 20 8222 8169 Postcode: 510730
Tel: +86 20 8213 9001 Fax: +86 20 8208 9909 Postcode: 510663



Test Report

HKTDR2020071080



Technical Documentation Review Report

Date: 23 Jul 2020

No. HKTDR2020071080

Page 1 of 4

SUNWAY (SHENZHEN) PRODUCTS LIMITED

101, 201, East of 301, 401 & 501, Building E, Jurouan Company, No.16 Yinkui Road, Kuisin Community, Kuichong Street, Dapeng New District, Shenzhen, 518119, China

The documentation was submitted by the client for the product as: **Disposable medical face mask (non-sterile)**

SGS Job No.
Model/Type
Manufacturer
Address of Manufacturer

GZHL2006034257MD
SW-0014
SUNWAY (SHENZHEN) PRODUCTS LIMITED
101, 201, East of 301, 401 & 501, Building E, Jurouan Company, No.16 Yinkui Road, Kuisin Community, Kuichong Street, Dapeng New District, Shenzhen, 518119, China

Country of Origin
Country of Destination
Date of Documentation Received
Review Period

CHINA
Europe
23 Jul 2020
23 Jul 2020 - 24 Jul 2020

Service Requested

Review the completeness of the Technical Documentation in accordance with the requirements of Annex VII of COUNCIL DIRECTIVE 93/42/EEC of 14 June 1993 concerning medical devices and its amendments as well as relevant harmonized standards.

Review Summary

Based on the submitted documentation, this is to conclude that non-compliance or missing information was not identified according to the requirements of Annex VII of COUNCIL DIRECTIVE 93/42/EEC of 14 June 1993 concerning medical devices and its amendments as well as relevant harmonized standards. Please note that no sample is provided for review and no testing is carried out in this service.


Date of Expiry

23 Jul 2021 (1 year)

DISCLAIMER - PLEASE READ

While all due care and skill were exercised in carrying out this review, SGS Hong Kong Ltd (SGS) accepts responsibility only for proven gross negligence. Considering that the situation surrounding COVID-19 is evolving, the findings provided in this report may change on a daily basis. This report relates only to the submitted documentation. The authenticity of the documentation reviewed by SGS, and the consistency of the product and the documentation is not covered. Conformance to all the regulatory requirements is the sole responsibility of the manufacturer including the production and quality control of the product(s). This report is not a legal document and cannot be used as such. This is not a legal interpretation of the law. Reliance should be placed on the wording of the legislation itself. SGS may have extracted from the compiled data specific criteria which are not intended to be the substitute of the relevant legislation and/or standards.

Signed for and on behalf of
SGS Hong Kong Ltd



Ivan CHAN
Vice President - R & D and Innovation

This document is issued by the Company subject to its General Conditions of Service printed separately available on request at www.sgs.com/Products/Forms-and-Conditions.aspx and for all other legal documents subject to Rules and Conditions for Technical Drawings at www.sgs.com/Products/Technical-Drawings.aspx

where it refers to the liability of safety, information and products shall remain the same. Any failure of this document to reflect the information contained herein reflects the Company's failure to the best of its reasonable skills and within the limits of its best endeavours. It does not constitute an offer of insurance or any other financial product or a recommendation from a financial adviser or any other financial product. This document is issued by the Company and its subsidiaries and affiliates and is not intended to be a substitute for the relevant legislation and/or standards. This document is issued by the Company and its subsidiaries and affiliates and is not intended to be a substitute for the relevant legislation and/or standards. This document is issued by the Company and its subsidiaries and affiliates and is not intended to be a substitute for the relevant legislation and/or standards.

Test Report


HKTDR2020071080

SGS

Technical Documentation Review Report
Date: 23 Jul 2020

No. HKTDR2020071080
Page 2 of 4

Product photo:



Product packaging and label artwork:



This document is issued by the Company, subject to the General Conditions of Sale and the specific conditions of sale, available on the website www.sgslab.com/hk/zh-hk/terms-and-conditions and the applicable General Conditions, Subjective Terms and Conditions for Business Situations or www.sgslab.com/hk/zh-hk/terms-and-conditions.

Notwithstanding to the limitation of liability, indemnification and jurisdiction clauses defined herein, the number of this document is applied for information purposes only. The Company's sole responsibility is in its Terms and Conditions and not extendable further to a transaction from execution of these rights and obligations under the transaction documents. This document cannot be reproduced, copied or full without prior written approval of the Company. Any reproduction, copying or further use of this document is prohibited and will be prosecuted by the legal action of the law.

Test Report

HKTDR2020071080

Technical Documentation Review Report

Date: 23 Jul 2020

No. HKTDR2020071080

Page 3 of 4

Review details

Document 001

Document Title	Test report	Recommendation on acceptance	Accepted
Document Number	20R000123MT		
Issuing party	GTTC		
Date of Issue	20-Jul-2020		
Description of Document	Test report according to EN 14683:2019+AC:2019		
Comment	The submitted Test Report was prepared according to EN 14683:2019+AC:2019.		

Document 002

Document Title	Technical Documentation	Recommendation on acceptance	Accepted
Document Number	SEW-CE001		
Issuing party	SUNWAY (SHENZHEN) PRODUCTS LIMITED		
Date of Issue	20-Jul-2020		
Description of Document	Technical Documentation prepared according to the Medical Device Directive 93/42/EEC.		
Comment	The submitted technical documentation was prepared according to the Annex VI of the Medical Device Directive 93/42/EEC as amended by 2007/47/EC.		

Document 003

Document Title	EC Declaration of Conformity	Recommendation on acceptance	Accepted
Document Number	SEW-CE001 APPENDIX H		
Issuing party	SUNWAY (SHENZHEN) PRODUCTS LIMITED		
Date of Issue	03-Jun-2020		
Description of Document	Declaration of Conformity (DoC) issued by SUNWAY (SHENZHEN) PRODUCTS LIMITED in order to fulfil the obligation according to the Medical Device Directive 93/42/EEC.		
Comment	Based on the submitted DoC, this is to conclude that non-compliance or missing information was not identified according to the requirements of Medical Device Directive 93/42/EEC.		

Test Report

HKTDR2020071080

SGS

Technical Documentation Review Report

Date: 23 Jul 2020

No. HKTDR2020071080

Page 4 of 4

Additional comments

According to Article 9 and Annex IX of COUNCIL DIRECTIVE 93/42/EEC of 14 June 1993 concerning medical devices, Disposable medical face mask (non-sterile) (SW-0614) according to EN14683:2019+AC:2019 "Medical face masks - Requirements and test methods" is considered as class 1 medical device.

According to Article 11 of the directive, conformity procedure as stated in Annex VII of the directive shall be followed.

According to Announcement No.12 (2020) of the Ministry of Commerce, People's Republic of China, as of April 28, 2020, exporting companies of SARS-CoV-2 testing reagents, medical face masks, medical protective suits, ventilators and infrared thermometers that have obtained certification or authorization from other countries shall submit an Export Declaration of Medical Supplies in writing together with customs declarations, as a warranty that the products are compliant with the quality standards and safety requirements of the importing countries (regions).

*** End of Report ***

FDA



Certification of FDA Registration Service

This document provides notification of the registration number assigned to the client's establishment.

Establishment: SUNWAY (SHENZHEN) PRODUCTS LIMITED
Address: E Building, Jiansuan company, 16# Yankui Road, Kuixin community, Kuichong street, Dapeng New District, Shenzhen, Guangdong, 518119, CHINA

Owner/Operator No.: 10064363
Registration No.: 3014331592

Listing No.	Premarket Submission No.	Product Codes	Device Name	Activities
D378991	Exempt	MSH	Respirator, surgical	Manufacturer
D378898	Exempt	LYU	ACCESSORY, SURGICAL APPAREL	Manufacturer
D403677	Exempt	FXD	SUIT, SURGICAL	Manufacturer
D403676	Exempt	DEA	Non-surgical isolation gown	Manufacturer
D403678	Exempt	KHA	MASK, SCAVENGING	Manufacturer
D403673	Exempt	FVF	CAP, SURGICAL	Manufacturer

FDA

D403675	Exempt	EYE	DRESS, SURGICAL	Manufacturer
D403674	Exempt	EXP	COVER, SHOE, OPERATING-ROOM	Manufacturer

Field Fu
Mr. Field Fu
SHENZHEN JOYANTECH CONSULTING CO., LTD
www.cefila.com

Certification No.: FDA 202003-005 Validity: March 25, 2020 - Dec. 31st, 2020

FDA

Business Trade

Name : *sunway*
(shenzhen)

Establishment

Registration or FEI

Number : 3014331592

[New Search⁶](#)

Establishment Name 7 8	Registration Number	Current Registration Yr
SUNWAY (SHENZHEN) PRODUCTS LIMITED⁹ CHINA	3014331592	2020
<ul style="list-style-type: none"> accessory, surgical apparel - Accessory,Surgical Apparel; SUN-V Blocklite/ Sunvia/ Sun Valley, Disposable Medical Protective Gowns, size 165/175/185/190/195; SUN-V Blocklite/ Sunvia/ Sun Valley, Disposable Protective Clothing, size 165/175/185/190/195; SUN-V Blocklite/ Sunvia/ Sun Valley, Medical Disposable Goggles¹⁰ 		Manufacturer
<ul style="list-style-type: none"> respirator, surgical - SUN-V Blocklite/ Sunvia/ Sun Valley, Disposable Medical Face Mask¹¹ 		Manufacturer
<ul style="list-style-type: none"> cap, surgical - Cap,Surgical¹² 		Manufacturer
<ul style="list-style-type: none"> cover, shoe, operating-room - Cover,Shoe,Operating-Room¹³ 		Manufacturer
<ul style="list-style-type: none"> dress, surgical - Dress,Surgical¹⁴ 		Manufacturer
<ul style="list-style-type: none"> non-surgical isolation gown - Non-Surgical Isolation Gown¹⁵ 		Manufacturer
<ul style="list-style-type: none"> suit, surgical - Suit,Surgical¹⁶ 		Manufacturer
<ul style="list-style-type: none"> mask, scavenging - Face mask; Protective mask KN95¹⁷ 		Manufacturer

FDA

FDA

1 result found for **Establishment** [New Search](#)
Registration or Business Trade Name : *sunway (shenzhen)*

Establishment Name	Registration Number	Current Registration Yr
SUNWAY (SHENZHEN) PRODUCTS LIMITED	CHINA No number listed	2020

- [Accessory, Surgical Apparel - SUN-V Blocklite/ Sunvia/ Sun Valley, Disposable Medical Protective Gowns, Size 165/175/185/190/195;](#) Manufacturer

[SUN-V Blocklite/ Sunvia/ Sun Valley, Disposable Protective Clothing, Size 165/175/185/190/195; SUN-V Blocklite/ Sunvia/ Sun Valley, Medical Disposable Goggles](#)

- [Respirator, Surgical - SUN-V Blocklite/ Sunvia/ Sun Valley, Disposable Medical Face Mask](#) Manufacturer

Can't find what you're looking for? [Try a new search](#)

FDA

FDA

[New Search](#) [Back To Search Results](#)

Proprietary Name: SUN-V Blocklite/ Sunvia/ Sun Valley, Disposable Medical Protective Gowns, size 165/175/185/190/195; SUN-V Blocklite/ Sunvia/ Sun Valley, Disposable Protective Clothing, size 165/175/185/190/195; SUN-V Blocklite/ Sunvia/ Sun Valley, Medical Disposable Goggles

Classification Name: ACCESSORY, SURGICAL APPAREL

Product Code: [LYU](#)

Device Class: 1

Regulation Number: [878.4040](#)

Medical Specialty: General & Plastic Surgery

Registered Establishment Name: [SUNWAY \(SHENZHEN\) PRODUCTS LIMITED](#)

Owner/Operator: [SUNWAY \(SHENZHEN\) PRODUCTS LIMITED](#)

Owner/Operator Number: 10064363

Establishment Operations: Manufacturer

-

-

Page Last Updated: 03/30/2020

Note: If you need help accessing information in different file formats, see [Instructions for Downloading Viewers and Players](#).

Language Assistance Available: [Español](#) | [繁體中文](#) | [Tiếng Việt](#) | [한국](#)
[한국](#) | [Tagalog](#) | [Русский](#) | [العربية](#) | [Kreyòl](#)
[Avisyen](#) | [Français](#) | [Polski](#) | [Português](#) | [Italiano](#) | [Deutsch](#) | [日本語](#) | [فارسی](#) | [English](#)

FDA

[New Search](#) [Back To Search Results](#)

Proprietary Name: SUN-V Blocklite/ Sunvia/ Sun Valley,

FDA

[SUN-V Blocklite/ Sunvia/ Sun Valley,
Disposable Protective Clothing, Size
165/175/185/190/195; SUN-V Blocklite/
Sunvia/ Sun Valley, Medical Disposable
Goggles](#)

- [Respirator, Surgical - SUN-V Blocklite/ Sunvia/
Sun Valley, Disposable Medical Face Mask](#) Manufacturer

Can't find what you're looking for? [Try a new search](#)

FDA

